

# Manual Supplement

Manual Title: 1587/1577 Users  
Part Number: 2401027  
Print Date: April 2005  
Revision/Date: 2, 6/09

Supplement Issue: **3**  
Issue Date: 5/13  
Page Count: 2

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1, 54524

On page 34, replace the **Ohms Measurement** table with the following:

### Ohms Measurement

Range	Resolution	Accuracy 1587 and 1587T <sup>[1]</sup> +(% of Rdg+Digits)	Accuracy 1577 <sup>[1]</sup> +(% of Rdg+Digits)
600.0 Ω	0.1 Ω	±(0.9 % + 2)	±(1.2 % + 2)
6.000 kΩ	0.001 kΩ		
60.00 kΩ	0.01 kΩ		
600.0 kΩ	0.1 KΩ		
6.000 MΩ	0.001 MΩ		
50.0 MΩ <sup>[2]</sup>	0.01 MΩ	±(1.5 % + 3)	±(2.0 % + 3)
[1] Accuracies apply from 0 to 100 % of range.			
[2] Up to 80 % relative humidity.			

## Change #2, 63834, 65441

On page 3, add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
CAT II	Measurement Category II is applicable to test and measuring circuits connected directly to utilization points (socket outlets and similar points) of the low-voltage MAINS installation.

CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
CAT IV	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.

On page 29, under **General Specifications**, remove the Shock, Electromagnetic Compatibility, Safety, Certifications, IP Rating and add:

- IP Rating** ..... IEC 60529: IP40
- Safety** ..... IEC 61010-1: 600 V CAT IV / 1000 V CAT III, Pollution Degree 2
- Electromagnetic Environment** ..... IEC 61326-1: Portable
- Electromagnetic Compatibility** ..... In an RF field of 3 V/m, accuracy = specified accuracy except in temperature accuracy = specified accuracy  $\pm 5^{\circ}\text{C}$  ( $9^{\circ}\text{F}$ ).  
 Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment) <sup>[1]</sup>

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.