

Supplement

Title:	373 Users	Supplement Issue:	4
Part Number:	3612805	Issue Date:	11/12
Print Date:	July 2010	Page Count:	2
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

Change #1, 55883

On page 15, under *DC Voltage*, add flag note [1] to Accuracy and add:

- [1] In an RF field greater than 1 V/m between 130 MHz and 350 MHz, add 65 counts to specified accuracy.



Change #2, 56248, 64117

On page 3, under **Warning**,

Measure a known voltage first to make sure that the Meter operates correctly.

Change #3

On page 6, add the following to the **Symbols** table:

	Conforms to CAN/CSA-C22.2 No. 61010-1, second edition, including Amendment 1.
	Conforms to relevant South Korean EMC Standards.

Change #4, 64117, 64124

On page 17, delete EMI, EMC and the Agency Approvals sections and add the following new sections:

EMC

Applies to use in Korea only.....Class A Equipment (Industrial Broadcasting & Communication Equipment) ^[1]

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

Safety.....IEC 61010-1: 300 V CAT IV / 600 V CAT III, Pollution Degree 2

EM Environment.....IEC 61326-1: Portable