

Manual Supplement

Manual Title: 724 Users
Print Date: February 2000
Revision/Date: 1, 8/03

Supplement Issue: **4**
Issue Date: 2/11
Page Count: 2

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 724
CD Rev. & Date: 2, 8/2003
CD PN: 1547849

Change #1

On page 41, add the following to the *Temperature, Thermocouples* table:

Type	Range	Measure and Source Accuracies (ITS-90)
XK	-200 to 800 °C	-200 to -100 ± 0.5 °C -100 to 800 ± 0.6 °C
BP	0 to 2500 °C	0 to 800 ± 1.2 °C 800 to 2500 ± 2.5 °C

Change #2

On page 16, under the *Note*, replace the first sentence with the following:

Available with V1.4 Firmware or greater.


Change #3, 41513

On page 43, under **General Specifications**, add the following:

Protection Class	Pollution Degree II
------------------	---------------------

Change #4

On page 7, Table 2, add the following:

	Conforms to relative Australian standards.
---	--

Change #5

On page 37, Table 7:

Change:

18	Test lead, red Test lead, black	688051 688066	1 1
----	------------------------------------	------------------	--------

To:

18	Fluke-7XX Test Lead Set	3397308	1
----	-------------------------	---------	---